

Catálogo: K051-2

CÁLCIO ARSENAZO III

Nº de testes: 500

Versão: 07/11/2018

Observações: Cód.: 00

Os reagentes são prontos para uso.

Name

Short Name

Sample Type  Analysis Mode  Unit  Decimals  No Replicates

Reaction Type  Report's Name

Reading Mode  Main Filter  Reference Filter

**Volumes** Sample  Vol. R1  Vol. R2

**Times** Reading 1:  (S)  Cycle

Reading 2:

**Blank Mode** Blank Type  Blank Replicate

**Calibration Mode**

Factor x

Experimental Calibrator Calibrator Replicates

Use Alternative Calibrator

Linearity Limit  Min Value  Max Value

Detection Limit  Normality

Predilution Factor

Predilution Mode  1/  Diluent

Automatic Repetition

Post Dilution Factor 1/

Increased X

**Calibration and Curve Values**

Number of Calibrators

N	Concentration	Factor
<input type="text" value="1"/>	<input type="text" value="#"/>	<input type="text" value="-"/>

Name  Lot

Expiration

Calibration Curve

Increasing  Decreasing

X-Asis:  Y-Asis:

Notas: A Bioclin recomenda o uso do Reagente Nº 2 (Padrão 10.0 mg/dL) ou do calibrador multiparâmetro de bioquímica Biocal – K072 para Calibração. Para avaliar a precisão e a exatidão das dosagens, recomendamos o uso dos soros controle Biocontrol N – K073 e Biocontrol P – K074.

Cada Laboratório Clínico deve possuir um programa interno de Controle de Qualidade.