

Catálogo: K049-6

TRANSAMINASE ALT (TGP) CINÉTICA

Nº de testes: 500

Versão: 14/08/2017

Observações: Cód.: 00

Os reagente são prontos para uso.

Name

Short Name

Sample Type	Analysis Mode	Unit	Decimals	No Replicates
<input type="text" value="Serum"/>	<input type="text" value="Birreag. Kinetic"/>	<input type="text" value="U/L"/>	<input type="text" value="2"/>	<input type="text" value="1"/>

Reaction Type	Report's Name
<input type="text" value="Decreasing"/>	<input type="text"/>

Reading Mode	Main Filter	Reference Filter
<input type="text" value="Monochromatic"/>	<input type="text" value="340"/>	<input type="text" value="-"/>

Volumes			Times		
Sample	Vol. R1	Vol. R2		(S)	Cycle
<input type="text" value="20 µL"/>	<input type="text" value="160 µL"/>	<input type="text" value="40 µL"/>	Reading 1:	<input type="text" value="360"/>	<input type="text" value="40"/>
			Reading 2:	<input type="text" value="495"/>	<input type="text" value="55"/>

Blank Mode	
Blank Type	<input type="text" value="Blank with Distilled Water"/>
Blank Replicate	<input type="text" value="1"/>

Calibration Mode	
<input type="radio"/> Factor	x <input type="text"/>
<input checked="" type="radio"/> Experimental Calibrator	Calibrator Replicates <input type="text" value="1"/>
<input type="radio"/> Use Alternative Calibrator	

Linearity Limit	<input type="text" value="400"/>		
Detection Limit	<input type="text" value="1.4"/>	Reference Ranges	
		Min Value	Max Value
		Normality	<input type="text" value="5.0"/> <input type="text" value="38.0"/>

<input type="radio"/> Predilution Factor	
Predilution Mode	Diluent
<input type="text"/>	1/ <input type="text"/> <input type="text"/>

<input type="checkbox"/> Automatic Repetition	
Post Dilution Factor	1/ <input type="text"/>
Increased	X <input type="text"/>

Calibration and Curve Values		
Number of Calibrators	<input type="text" value="1"/>	
N	Concentration	Factor
<input type="text" value="1"/>	<input type="text" value="#"/>	<input type="text" value="-"/>
Name	<input type="text" value="Biocal - K072"/>	
Lot	<input type="text" value="-"/>	
Expiration	<input type="text" value="-"/>	

Calibration Curve	
<input checked="" type="radio"/> Increasing	<input type="radio"/> Decreasing
<input type="text"/>	
X-Asis:	Y-Asis:
<input type="text"/>	<input type="text"/>

Notas: A Bioclin recomenda o uso do calibrador multiparâmetro de bioquímica Biocal – K072 para Calibração. Para avaliar a precisão e a exatidão das dosagens, recomendamos o uso dos soros controle Biocontrol N – K073 e Biocontrol P – K074.

Cada Laboratório Clínico deve possuir um programa interno de Controle de Qualidade.